

## CERTIFICATE OF ANALYSIS

# Buffer Solution

# 12.45 pH

Item code:	DC112620 (05/1/107-NR)
Batch code:	11574
Exp. Date:	09/2017
Manufactured:	09/2015
Tolerance:	+/- 0.05 pH
Nominal pH-value:	12.45 pH (at 25°C)
Actual pH-value:	12.45 pH (at 25°C)
Secondary Ref. Material:	Calciumhydroxide
Ref. Material LOT-No.:	142111A

### Standardization method:

Actual pH-value as stated above was obtained using a combination glass electrode according to DIN 19268.

Calibration is traceable to national and international standards, PBT and NIST validated by key comparisons.