

CERTIFICATE OF ANALYSIS

Buffer Solution

9.18 pH

Item code:	DC109306 (05/1/107-NR)
Batch code:	11624
Exp. Date:	02/2018
Manufactured:	11/2015
Tolerance:	+/- 0.02 pH
Nominal pH-value:	9.18 pH (at 25°C)
Actual pH-value:	9.18 pH (at 25°C)
Secondary Ref. Material:	Sodiumtetraborate
Ref. Material LOT-No.:	A0264508321

Standardization method:

Actual pH-value as stated above was obtained using a combination glass electrode according to DIN 19268.

Calibration is traceable to national and international standards, PBT and NIST validated by key comparisons.